
Western Regional Emergency Medical Advisory Committee (WREMAC)

Title: WREMAC Cardiac Monitor/Defibrillator Specifications for All New Equipment Purchases	Effective Date: January 4, 2008 Page 1 of 2
Policy # 1997-2	Revised: January 2008

Policy	All cardiac monitors / defibrillator equipment used in the WREMAC region shall be compliant with the specifications in the attached procedure.
Procedure	Follow attached guidelines
Reference	Western Regional Emergency Medical Advisory Committee (WREMAC) March, 2008 Minutes (approval)

PROCEDURE:	Automated	PAD/BLSFR/EMT-B	EMT-I	EMT-CC/P
	Voice Prompts	Yes	Yes	Optional
	Visual Prompts	Yes	Yes	Optional
	Hands-Free Defibrillation	Yes	Yes	Optional
	Ability to Print Code Summary for Receiving Hospital Within 24 Hours	Optional	Optional	Yes
	Ability to Print Real-time Rhythm Strip	Optional (Device option may be available to CC/P and Credentialed I's only)	Optional (Device option may be available to CC/P and Credentialed I's only)	Yes
	Screen/Display to Monitor Rhythm	Optional (Device option may be available to CC/P and credentialed I's only)	Optional (Device option may be available to CC/P and credentialed I's only)	Yes
	Manual Operation (Adult & Pediatric)	Optional (Device option may be available to CC/P only and must be controlled by lockout/password)	Optional (Device option may be available to CC/P only and must be controlled by lockout/password)	Yes
	Synchronized Cardioversion (Adult & Pediatric)	Optional (Device option may be available to CC/P only and must be controlled by lockout/password)	Optional (Device option may be available to CC/P only and must be controlled by lockout/password)	Yes
	Pacing (Adult & Pediatric)	Optional (Device option may be available to CC/P only and must be controlled by lockout/password)	Optional (Device option may be available to CC/P only and must be controlled by lockout/password)	Yes
	Defibrillation (Adult & Pediatric)	Yes	Yes	Yes
	Bi-Phasic Capability	Yes	Yes	Yes
	Waveform Capnography (Adult & Pediatric)	Optional (Device option may be available to I/CC/P only)	Yes	Yes
	12 Lead Monitoring Capability	Optional (Device option may be available to CC/P only and controlled by lockout/password)	Optional (Device option may be available to CC/P only and controlled by lockout/password)	Yes
	Ability to transmit 12-lead EKG	Optional (Device option may be available to CC/P only and controlled by lockout/password)	Optional (Device option may be available to CC/P only and controlled by lockout/password)	Recommended